

DATA SHEET: DSN538

PRODUCT SPECIFICATION

Product name	Latex Examination Gloves
Range	Healthgard
Classification	
Marking	Product name, Premier Logo, Size, Code Ref., Lot Number, Expiry date and CE
	Mark
Material	NR latex, zinc oxide, sulphur, hindered phenol, titanium dioxide
Packaging	Pouch, dispenser, carton
Sterilization	non sterile

Primary use	Ideal for working with fluids or wet glassware	

Product code	Colour	size	Quantity/inner box	Quantity/inner box
PSCR0036	Natural	Extra Small	100	1,000
PSCR0037	Natural	Small	100	1,000
PSCR0038	Natural	Medium	100	1,000
PSCR0039	Natural	Large	100	1,000
PSCR0040	Natural	Extra Large	100	1,000

dimension	Minimum Thickness	Minimum Length	Palm Width of Gloves
PSCR0036	0.15 mm	240 mm	76 +/-3mm
PSCR0037	0.15 mm	240 mm	84 +/-3mm
PSCR0038	0.15 mm	240 mm	94 +/-3mm
PSCR0039	0.15 mm	240 mm	105 +/-3mm
PSCR0040	0.15 mm	240 mm	113 +/-3mm

Quality certification	ISO 13485		
Conformity to standards	BS.EN455 Parts 1&2		
Storage conditions	Store in a cool dry place away from heat and direct sunlight		
Shelf life	3 Years- if stored as recommended above		
Additional information	Presige latex examination gloves have an enhanced textured surface to provide excellent touch sensitivity and grip, which is ideal when working with fluids, instruments or wet glassware. Fully micro textured gloves, single chlorinated, powder free, conforms to BS EN 455 standards, single use		

Medical Device and Reporting System:

According to the official criteria of MDD, incidents caused by this device must be reported immediately to our offices.

The information provided within this Data Sheet is correct as of last revision but it is subject to change without notice.

Shermond T: +44 (0) 1530 278 111 W: <u>www.shermond.com</u> Unit 6, Delta Park Industrial Estate Millmarsh Lane, London Enfield, EN3 7QJ

Shermond is a trading division of Bunzl Retail & Healthcare Supplies



DATA SHEET: DSN538

REVISON:	2.0	DATE	03/01/2013	AUTHOR	H Cocking
NOTES:	Updated to new format and periodic review				
REVISON:	3.0	DATE	01/23/2017	AUTHOR	R.Zhang
NOTES:	Periodic review, DSN format and Shermond logo updated.				

The information provided within this Data Sheet is correct as of last revision but it is subject to change without notice.