

PRODUCT SPECIFICATION

Product name	General Purpose Beaker
Range	Universal
Classification	Medical Device - Class 1
Marking	Product name, Universal Logo, Size, Code ref, lot number, expiry date and CE mark
Material	Polypropylene
Packaging	carton

Primary use	Polypropylene disposable general purpose beaker for collection and evaluation of samples and distribution of medicines
--------------------	--

Product code	Dimension
UNH7-8005	300ml

Quality certification	ISO 13485
Storage conditions	Store in a cool dry place away from heat and direct sunlight
Shelf life	5 Years- if stored as recommended above
Additional information	These can be autoclaved (121°C 20 Minutes), latex free. They are not intended to measure liquids

*Medical Device and Reporting System:
According to the official criteria of MDD, incidents caused by this device must be reported immediately to our offices.*

REVISION:	1.0	DATE	10/03/2013	AUTHOR	H Cocking
NOTES:	Updated to new format and periodic review				
REVISION:	2.0	DATE	02/17/2017	AUTHOR	R.Zhang
NOTES:	Periodic review, DSN format and Shermond logo updated.				

The information provided within this Data Sheet is correct as of last revision but it is subject to change without notice.

Shermond T: +44 (0) 1530 278 111 W: www.shermond.com	Unit 6, Delta Park Industrial Estate Millmarsh Lane, London Enfield, EN3 7QJ
--	--